

Requirements for Lead Testing*

Oklahoma

* This information is subject to change

Licensure

Facility must: Hold a CLIA Certificate of Waiver or higher.

1. Complete the application for a CLIA Certificate of Waiver (<http://www.cms.hhs.gov/cmsforms/downloads/cms116.pdf>)
2. Mail completed applications to:
OKLAHOMA STATE DEPARTMENT OF HEALTH
Protective Health Services
Medical Facilities
1000 NE 10TH Street
Oklahoma City, OK 73117

For questions or more information, please call (405) 271-6576 or e-mail medicalfacilities@health.ok.gov.

Patient Testing

Training Tools: www.LeadCare2.com/training

Facility must: Confirm capillary blood lead test results greater than or equal to 5 µg/dL with a venous sample by a reference lab.

Quality Control

Facility must: Run two levels of Quality Control according to the manufacturer's instructions, which are:

1. Each new lot of test kits.
2. Each new shipment of materials even if it's the same lot previously received.
3. Each new operator (i.e. operator who has not performed the test recently).
4. Monthly, as a check on continued storage conditions.
5. When problems (storage, operator, instrument, or other) are suspected or identified.
6. If otherwise required by your laboratory's standard QC procedures.

Elevated Blood Lead Levels

Facility Must:

1. Report results of all blood lead levels greater than or equal to 5 µg/dL weekly (if possible daily)
2. Report results of all blood lead levels greater than or equal to 20 µg/dL within 24 hours of date of analysis

Result Reporting

Reporting Solutions by State: www.LeadCare2.com/reporting

Facility must:

1. Report ALL blood lead results, monthly (if possible, weekly), to:
Oklahoma Childhood Lead Poisoning Prevention Program
1000 NE 10th Street, Room 711
Oklahoma City, OK 73117-1299
Phone: (405) 271-6617
Fax: (405) 271-4971
Email: oklppp@health.ok.gov
2. Report via phone, fax, email with the OCLPPP reporting spreadsheet attached, or electronic data transmission (LeadCare® Report preferred).
 - o Go to www.leadcare2.com/reporting/ to download a FREE copy of LeadCare® Report Software.

****Please notify Magellan or the Oklahoma Childhood Lead Poisoning Prevention Program if the LeadCare® II instrument is transferred to another facility. This will ensure that the appropriate reporting education is conveyed to the end-user.**

Proficiency Testing

Recommended: To monitor the quality of your blood lead testing program.

Contact Wisconsin State Laboratory of Hygiene (WSLH) proficiency program for more information about their program at (800) 462-5261 or go to <http://www.slh.wisc.edu/proficiency/>

Still Have Questions?

- Call LeadCare Product Support at (800) 275-0102
- Call the Oklahoma CLPPP at (405) 271-6617